

IMMUNOASSAY SPECIALITY I - LEVEL 2 (IA SPECIALITY I LEV 2)

CAT. NO. IAS3114 **LOT NO.** 2421EC
SIZE 5 x 2 ml **EXPIRY:** 2026-04-28
GTIN: 05055273207309

INTENDED USE

This product is intended for *in vitro* diagnostic use, in the quality control of the accuracy of Immunoassays on clinical chemistry systems. This material can be used to monitor the control of accuracy or the control of reproducibility of immunoassays.

SAFETY PRECAUTIONS AND WARNINGS

For *in vitro* diagnostic use only. Do not pipette by mouth. Exercise the normal precautions required for handling laboratory reagents.

Human source material, from which this product has been derived, has been tested at donor level for the Human Immunodeficiency Virus (HIV 1, HIV 2) antibody, Hepatitis B Surface Antigen (HbsAg), and Hepatitis C Virus (HCV) antibody and found to be NON-REACTIVE. FDA approved methods have been used to conduct these tests.

However, since no method can offer complete assurance as to the absence of infectious agents, this material and all patient samples should be handled as though capable of transmitting infectious diseases and disposed of accordingly.

Health and Safety Data Sheets are available on request.

STORAGE AND STABILITY

Unopened: Unreconstituted control is stable up to the expiry date shown on the side of each individual bottle.

Opened: Once reconstituted, the components of the control are stable for 5 days at +2°C to +8°C and 28 days at -20°C when frozen once. Except:

Analyte	+2°C to +8°C	Aliquoted and frozen once at -20°C
Anti-TG	3 days	-
C-Peptide	1 day	-
IGF-I	8 hours	-
Osteocalcin	4 hours	-
Parathyroid Hormone (PTH)	4 hours	14 days
Procalcitonin	1 day	-

Only the required amount of product should be removed. After use, any residual product should NOT BE RETURNED to the original vial.

Bacterial contamination of the reconstituted serum will cause reductions in the stability of many components. If bacterial contamination is suspected, the vial should be discarded and a fresh vial reconstituted.

PREPARATION

The Immunoassay Speciality I Control is supplied lyophilised.

- Carefully reconstitute each vial of lyophilised serum with exactly 2 ml of distilled water at +15°C to +25°C. Close the bottle and allow to stand for 30 minutes before use. Ensure contents are completely dissolved by swirling gently. Avoid formation of foam. Do not shake.
- Refer to the Control section of the individual analyser application.
- Refrigerate any unused material. Prior to reuse, mix contents thoroughly.

MATERIALS PROVIDED

Immunoassay Speciality I - Level 2 5 x 2 ml

MATERIAL REQUIRED BUT NOT PROVIDED

Volumetric pipette

VALUE ASSIGNMENT

Due to the variation caused by test equipment, test reagents and laboratory technique, the quoted ranges are provided for guidance. It is recommended that these ranges are used until each laboratory has established its own ranges, based on individual laboratory requirements.

Each batch of Immunoassay Speciality I Control is submitted to a number of reference laboratories and values are assigned from a consensus of results obtained by these laboratories, using a unique statistical analysis. With each batch, a control range is provided for individual parameters and each parameter method. The control range is equivalent to the assigned mean ± 2 S.D.

If a method is unavailable, contact Randox Laboratories - Technical Services, Northern Ireland, tel: +44 (0) 28 9445 1070 or email Technical.Services@randox.com.

EC	REP
----	-----

Randox Teoranta, Meenmore,
Dungloe, Donegal,
F94 TV06, Ireland

30 Jan '24 me

IMMUNOASSAY SPECIALITY I - LEVEL 2 (IA SPECIALITY I LEV 2)

Cat. No. IAS3114 Lot. No. 2421EC

Size 5 x 2 ml Expiry 2026-04-28

Range					
Analyte	unit	Target	low	high	methods
1-25 (OH) ₂ Vitamin D	pmol/l	184	129	239	Diasorin Liaison XL
25-OH Vitamin D	nmol/l	54.2	40.7	67.8	Diasorin Liaison
	nmol/l	52.7	39.5	65.9	Abbott Architect (3L52)
	nmol/l	76.3	57.2	95.4	Siemens Centaur XP/XPT/Classic
	nmol/l	73.7	55.3	92.1	Vitros ECI
	nmol/l	62.1	46.6	77.6	BioMerieux Vidas
	nmol/l	49.0	36.8	61.3	Beckman Dxl 600/800
	nmol/l	52.0	39.0	65.0	Abbott Architect (5P02)/ Alinity (8P45)
	nmol/l	54.4	40.8	68.0	Diasorin Liaison XL
	nmol/l	49.1	36.8	61.4	Roche Vitamin D Total II
	nmol/l	75.9	56.9	94.9	Siemens Atellica IM
	nmol/l	45.9	34.4	57.4	Roche Vitamin D Total II e801
	nmol/l	50.4	37.8	63.0	Roche Vitamin D Total III
Anti TG	kU/l	91.7	68.8	115	Siemens Immulite 2000/2500
	kU/l	127	95.3	159	Abbott Architect / Alinity
	kU/l	510	383	638	Roche Cobas 4000/E411
	kU/l	546	410	683	Roche Cobas e601/602
	kU/l	199	149	249	Beckman Dxl 600/800
	kU/l	132	99.0	165	Siemens Centaur XP/XPT/Classic
	kU/l	219	164	274	Beckman Access/LXi725
	kU/l	207	155	259	BioMerieux Vidas
	kU/l	140	105	175	Siemens Atellica IM
	kU/l	484	363	605	Roche Cobas e402/e801
	kU/l	143	107	179	Siemens Atellica IM aTgII
	kU/l	136	102	170	Siemens Centaur XP/XPT aTgII
Anti TPO	kU/l	114	85.5	143	Abbott Architect / Alinity
	kU/l	127	95.3	159	Siemens Immulite 2000/2500
	kU/l	85.2	63.9	107	Roche Cobas 4000/E411
	kU/l	81.7	61.3	102	Roche Cobas c501/502 e601/602
	kU/l	67.0	50.3	83.8	Beckman Dxl 600/800
	kU/l	437	328	546	Siemens Centaur XP/XPT/Classic
	kU/l	74.4	55.8	93.0	Beckman Access/LXi725
	kU/l	102	76.5	128	BioMerieux Vidas
	kU/l	109	81.8	136	Roche Cobas e402/e801
	kU/l	419	314	524	Siemens Atellica IM
C-Peptide	nmol/l	0.890	0.668	1.110	Siemens Centaur XP/XPT/Classic
	ng/ml	2.69	2.02	3.36	
	nmol/l	0.970	0.728	1.210	Roche Cobas e601/602
	ng/ml	2.93	2.20	3.66	
	nmol/l	0.870	0.653	1.090	Abbott Architect / Alinity
	ng/ml	2.63	1.97	3.29	
	nmol/l	0.800	0.600	1.000	Siemens Immulite 2000/2500
	ng/ml	2.42	1.81	3.03	

IMMUNOASSAY SPECIALITY I - LEVEL 2 (IA SPECIALITY I LEV 2)

Cat. No. IAS3114 Lot. No. 2421EC

Size 5 x 2 ml Expiry 2026-04-28

Range					
Analyte	unit	Target	low	high	methods
C-Peptide	nmol/l	0.960	0.720	1.200	Roche Cobas 4000/E411
	ng/ml	2.90	2.17	3.63	
	nmol/l	1.140	0.855	1.430	Diasorin Liaison XL
	ng/ml	3.44	2.58	4.30	
	nmol/l	0.960	0.720	1.200	Roche Cobas e402/e801
	ng/ml	2.90	2.17	3.63	
	nmol/l	0.770	0.578	0.963	Siemens Atellica IM
	ng/ml	2.32	1.74	2.90	
IGF 1	µg/l	212	159	265	Siemens Immulite 2000/2500
	µg/l	64.8	48.6	81.0	Roche Cobas 4000/E411
	µg/l	70.9	53.2	88.6	Roche Cobas 6000/8000
	µg/l	63.9	47.9	79.9	Roche Elecsys
Insulin	mU/l	8.74	6.56	10.9	Abbott Architect
	µU/ml	8.74	6.56	10.9	
	mU/l	10.2	7.65	12.8	Roche Cobas 6000/8000
	µU/ml	10.2	7.65	12.8	
	mU/l	11.4	8.55	14.3	Siemens Centaur XP/XPT/Classic
	µU/ml	11.4	8.55	14.3	
	mU/l	9.19	6.89	11.5	Beckman Dxl 600/800
	µU/ml	9.19	6.89	11.5	
	mU/l	10.7	8.03	13.4	Roche Cobas 4000/E411
	µU/ml	10.7	8.03	13.4	
	mU/l	10.5	7.88	13.1	Siemens Atellica IM
	µU/ml	10.5	7.88	13.1	
	mU/l	8.10	6.08	10.1	Abbott Alinity i
	µU/ml	8.10	6.08	10.1	
Osteocalcin	µg/l	36.2	27.2	45.3	Siemens Immulite 2000/2500
	µg/l	78.6	59.0	98.3	Roche Cobas 6000/8000
	µg/l	75.0	56.3	93.8	Roche Cobas 4000/E411
Parathyroid Hormone (PTH)	pmol/l	42.4	31.8	53.0	Abbott Architect / Alinity
	pg/ml	403	302	504	
	pmol/l	31.3	23.5	39.1	Beckman Dxl 600/800
	pg/ml	297	223	371	
	pmol/l	27.1	20.3	33.9	Roche Cobas e601/602 PTH
	pg/ml	257	193	321	
	pmol/l	29.1	21.8	36.4	Roche Cobas 4000/e411 PTH/PTH STAT
	pg/ml	276	207	345	
	pmol/l	38.1	28.6	47.6	Siemens Centaur XP/XPT/Classic
	pg/ml	362	272	452	
	pmol/l	26.8	20.1	33.5	Roche Cobas e601/602 PTH (1-84)
	pg/ml	255	191	319	
	pmol/l	29.2	21.9	36.5	Roche Cobas 4000/E411 PTH (1-84)
	pg/ml	277	208	346	
	pmol/l	37.8	28.4	47.3	Siemens Atellica IM
	pg/ml	359	270	448	

IMMUNOASSAY SPECIALITY I - LEVEL 2 (IA SPECIALITY I LEV 2)

Cat. No. IAS3114 Lot. No. 2421EC

Size 5 x 2 ml Expiry 2026-04-28

Range					
Analyte	unit	Target	low	high	methods
Parathyroid Hormone (PTH)	pmol/l	26.1	19.6	32.6	Roche Cobas e801 PTH
	pg/ml	248	186	310	
	pmol/l	25.5	19.1	31.9	Roche Cobas e801 PTH (1-84)
	pg/ml	242	181	303	
	pmol/l	24.9	18.7	31.1	Roche Cobas e801 PTH STAT
	pg/ml	236	178	294	
Procalcitonin	pmol/l	25.3	19.0	31.6	Roche Cobas e601/602 PTH STAT
	pg/ml	240	180	300	
	µg/l	1.76	1.32	2.20	Brahms Kryptor
	µg/l	4.67	3.50	5.84	BioMerieux Vidas
	µg/l	2.77	2.08	3.46	Roche Elecsys/Cobas/Modular
	µg/l	4.01	3.01	5.01	Siemens Centaur XP/XPT/Classic
	µg/l	2.74	2.06	3.43	SNIBE Maglumi Analysers
	µg/l	3.88	2.91	4.85	Siemens Centaur CP 10378883
	µg/l	2.69	2.02	3.36	Abbott Architect / Alinity Brahms PCT
	µg/l	2.72	2.04	3.40	Roche Cobas e402/e801
	µg/l	3.98	2.99	4.98	Siemens Atellica IM 10995651
	µg/l	3.63	2.72	4.54	Beckman Access PCT
	µg/l	3.93	2.95	4.91	Siemens Centaur XP/XPT 11202697
	µg/l	4.13	3.10	5.16	Siemens Centaur CP 11202697
	µg/l	4.05	3.04	5.06	Siemens Atellica IM 11202699

IMMUNOASSAY SPECIALITY I - LEVEL 2 (IA SPECIALITY I LEV 2)

CAT. NO. IAS3114 **LOT NO.** 2421EC
SIZE: 5 x 2 ml **EXPIRY:** 2026-04-28
GTIN: 05055273207309

INTENDED USE

This product is intended for *in vitro* diagnostic use, in the quality control of the accuracy of Immunoassays on clinical chemistry systems. This material can be used to monitor the control of accuracy or the control of reproducibility of immunoassays.

SAFETY PRECAUTIONS AND WARNINGS

For *in vitro* diagnostic use only. Do not pipette by mouth. Exercise the normal precautions required for handling laboratory reagents.

Human source material, from which this product has been derived, has been tested at donor level for the Human Immunodeficiency Virus (HIV 1, HIV 2) antibody, Hepatitis B Surface Antigen (HbsAg), and Hepatitis C Virus (HCV) antibody and found to be NON-REACTIVE. FDA approved methods have been used to conduct these tests.

However, since no method can offer complete assurance as to the absence of infectious agents, this material and all patient samples should be handled as though capable of transmitting infectious diseases and disposed of accordingly.

Health and Safety Data Sheets are available on request.

STORAGE AND STABILITY

Unopened: Unreconstituted control is stable up to the expiry date shown on the side of each individual bottle.

Opened: Once reconstituted, the components of the control are stable for 5 days at +2°C to +8°C and 28 days at -20°C when frozen once. Except:

Analyte	+2°C to +8°C	Aliquoted and frozen once at -20°C
Anti-TG	3 days	-
C-Peptide	1 day	-
IGF-I	8 hours	-
Osteocalcin	4 hours	-
Parathyroid Hormone (PTH)	4 hours	14 days
Procalcitonin	1 day	-

Only the required amount of product should be removed. After use, any residual product should NOT BE RETURNED to the original vial.

Bacterial contamination of the reconstituted serum will cause reductions in the stability of many components. If bacterial contamination is suspected, the vial should be discarded and a fresh vial reconstituted.

PREPARATION

The Immunoassay Speciality I Control is supplied lyophilised.

- Carefully reconstitute each vial of lyophilised serum with exactly 2 ml of distilled water at +15°C to +25°C. Close the bottle and allow to stand for 30 minutes before use. Ensure contents are completely dissolved by swirling gently. Avoid formation of foam. Do not shake.
- Refer to the Control section of the individual analyser application.
- Refrigerate any unused material. Prior to reuse, mix contents thoroughly.

MATERIALS PROVIDED

Immunoassay Speciality I - Level 2 5 x 2 ml

MATERIAL REQUIRED BUT NOT PROVIDED

Volumetric pipette

VALUE ASSIGNMENT

Due to the variation caused by test equipment, test reagents and laboratory technique, the quoted ranges are provided for guidance. It is recommended that these ranges are used until each laboratory has established its own ranges, based on individual laboratory requirements.

Each batch of Immunoassay Speciality I Control is submitted to a number of reference laboratories and values are assigned from a consensus of results obtained by these laboratories, using a unique statistical analysis. With each batch, a control range is provided for individual parameters and each parameter method. The control range is equivalent to the assigned mean ± 2 S.D.

If a method is unavailable, contact Randox Laboratories - Technical Services, Northern Ireland, tel: +44 (0) 28 9445 1070 or email Technical.Services@randox.com.



Randox Teoranta, Meenmore,
Dungloe, Donegal,
F94 TV06, Ireland

(S). Rev. 16 Oct '24 me

IMMUNOASSAY SPECIALITY I LEVEL 2 (IA SPECIALITY I LEV 2)

Cat. No. IAS3114 Lot No. 2421EC Size: 5 x 2ml Expiry: 2026-04-28

Analyte	unit	Target	Range		methods
			low	high	
25-OH Vitamin D	nmol/l	55.8	41.9	69.7	Beckman Access / Access 2
	nmol/l	49.0	36.8	61.3	Beckman Dxl600/800
Anti TG	kU/l	219	164	274	Beckman Access / Access 2
	kU/l	199	149	249	Beckman Dxl600/800
Anti TPO	kU/l	74.4	55.8	93.0	Beckman Access / Access 2
	kU/l	67.0	50.3	83.8	Beckman Dxl600/800
C-Peptide	nmol/l	0.576	0.432	0.720	Beckman Access
	ng/ml	1.74	1.31	2.17	
Insulin	mU/l	9.31	6.98	11.6	Beckman Access / Access 2
	mU/l	9.19	6.89	11.5	Beckman Dxl600/800
Parathyroid Hormone (PTH)	pmol/l	30.3	22.7	37.9	Beckman Access / Access 2
	pg/ml	288	216	360	
	pmol/l	31.3	23.5	39.1	Beckman Dxl600/800
	pg/ml	297	223	371	
Procalcitonin	µg/l	3.63	2.72	4.54	Beckman Access / Access 2
	µg/l	3.54	2.66	4.43	Beckman Dxl600/800